

**EMSL Analytical, Inc.**

200 Route 130, Cinnaminson, NJ, 08077
Telephone: 856-858-4800 Fax:cs@emsl.com
EMSL-CIN-01

EMSL Order ID: 012517208
LIMS Reference ID: AD17208
EMSL Customer ID: EPIC62

Attention: James Eberts

Epic Environmental Services, LLC [EPIC62]
80 Fork Bridge Road
Pittsgrove, NJ 08318
(856) 205-1077
jeberts@epicenviro.com

Project Name:

Paulsboro HS - LIW Additional

Customer PO:

24-2339C

EMSL Sales Rep:

Justin Monturano

Received:

04/16/2025 16:07

Reported:

04/30/2025 16:52

Analytical Results

Analyte	Result	Q	DF	RL	Units	Prepared Date/Time	Analyzed Date/Time	Analyst Initials	Prep /Analytical Method
Sample: PHS-FB Lims Reference ID: AD17208-01 Matrix: Drinking Water Sampled: 04/15/25 00:00:00									
Metals									
Lead	ND		1	1.00	µg/L	04/22/25 12:02	04/24/25 18:53	JW1	EPA 200.8 (DA)/EPA 200.8
Sample: PHS-WC-Glock Lims Reference ID: AD17208-02 Matrix: Drinking Water Sampled: 04/15/25 06:09:00									
Metals									
Lead	10.7		1	1.00	µg/L	04/22/25 12:02	04/24/25 18:59	JW1	EPA 200.8 (DA)/EPA 200.8
Sample: PHS-BF-Glock Lims Reference ID: AD17208-03 Matrix: Drinking Water Sampled: 04/15/25 06:10:00									
Metals									
Lead	12.8		1	1.00	µg/L	04/22/25 12:02	04/24/25 19:00	JW1	EPA 200.8 (DA)/EPA 200.8
Sample: PHS-WC-Admin Lims Reference ID: AD17208-04 Matrix: Drinking Water Sampled: 04/15/25 06:15:00									
Metals									
Lead	ND		1	1.00	µg/L	04/22/25 12:02	04/24/25 19:06	JW1	EPA 200.8 (DA)/EPA 200.8
Sample: PHS-BF1-Admin Lims Reference ID: AD17208-05 Matrix: Drinking Water Sampled: 04/15/25 06:16:00									
Metals									
Lead	ND		1	1.00	µg/L	04/22/25 12:02	04/24/25 19:08	JW1	EPA 200.8 (DA)/EPA 200.8
Sample: PHS-HB-OUT Lims Reference ID: AD17208-06 Matrix: Drinking Water Sampled: 04/15/25 06:21:00									
Metals									
Lead	5.25		1	1.00	µg/L	04/22/25 12:02	04/24/25 19:10	JW1	EPA 200.8 (DA)/EPA 200.8
Sample: PHS-WC-Garage Lims Reference ID: AD17208-08 Matrix: Drinking Water Sampled: 04/15/25 06:25:00									
Metals									
Lead	ND		1	1.00	µg/L	04/22/25 12:02	04/24/25 19:12	JW1	EPA 200.8 (DA)/EPA 200.8

**EMSL Analytical, Inc.**

200 Route 130, Cinnaminson, NJ, 08077
Telephone: 856-858-4800 Fax:cs@emsl.com
EMSL-CIN-01

EMSL Order ID: 012517208**LIMS Reference ID:** AD17208**EMSL Customer ID:** EPIC62**Attention:** James Eberts

Epic Environmental Services, LLC [EPIC62]
80 Fork Bridge Road
Pittsgrove, NJ 08318
(856) 205-1077
jeberts@epicenviro.com

Project Name:

Paulsboro HS - LIW Additional

Customer PO:

24-2339C

EMSL Sales Rep:

Justin Monturano

Received:

04/16/2025 16:07

Reported:

04/30/2025 16:52

Certified Analyses included in this Report

Analyte	Certifications
<i>EPA 200.8 in Drinking Water</i>	
Lead	NJDEP

List of Certifications

Code	Description	Number	Expires
PADEP	Pennsylvania Department of Environmental Protection	2845.25	11/30/2025
NYSDOH	New York State Department of Health ELAP	10872	04/01/2025
NJDEP	New Jersey Department of Environmental Protection	03036	06/30/2025
MADEP	Massachusetts Department of Environmental Protection	M-NJ337	06/30/2025
CTDPH	Connecticut Department of Public Health	PH-0270	06/23/2026
California ELAP	California Water Boards	1877	06/30/2025
AIHA LAP	EMSL Analytical, Inc. Cinnaminson, NJ AIHA-LAP, LLC-ELLAP Accredited	100194	05/01/2025
A2LA	A2LA Environmental Certificate	2845.01	07/31/2026

Please see the specific Field of Testing (FOT) on www.emsl.com <<http://www.emsl.com>> for a complete listing of parameters for which EMSL is certified.

**EMSL Analytical, Inc.**

200 Route 130, Cinnaminson, NJ, 08077
Telephone: 856-858-4800 Fax: cs@emsl.com
EMSL-CIN-01

EMSL Order ID: 012517208**LIMS Reference ID:** AD17208**EMSL Customer ID:** EPIC62**Attention:** James Eberts

Epic Environmental Services, LLC [EPIC62]
80 Fork Bridge Road
Pittsgrove, NJ 08318
(856) 205-1077
jeberts@epicenviro.com

Project Name:

Paulsboro HS - LIW Additional

Customer PO:

24-2339C

EMSL Sales Rep:

Justin Monturano

Received:

04/16/2025 16:07

Reported:

04/30/2025 16:52

Notes and Definitions

Item	Definition
(Dig)	For metals analysis, sample was digested.
[2C]	Reported from the second channel in dual column analysis.
DA	Direct Analysis
DF	Dilution Factor
MDL	Method Detection Limit.
ND	Analyte was NOT DETECTED at or above the detection limit.
NR	Spike/Surrogate showed no recovery.
Q	Qualifier
RCS	Respirable Crystalline Silica
RL	Reporting Limit
Wet	Sample is not dry weight corrected.

Measurement of uncertainty and any applicable definitions of method modifications are available upon request. Per EPA NLLAP policy, sample results are not blank corrected.

Owen McKenna Laboratory Manager or other approved signatory

EMSL maintains liability limited to cost of analysis. Interpretation and use of test results are the responsibility of the client. This report relates only to the samples reported above, and may not be reproduced, except in full, without written approval by EMSL. EMSL bears no responsibility for sample collection activities or analytical method limitations. The report reflects the samples as received. Results are generated from the field sampling data (sampling volumes and areas, locations, etc.) provided by the client on the Chain of Custody. Samples are within quality control criteria and met method specifications unless otherwise noted.



EMSL ANALYTICAL, INC.
LABORATORY PRODUCT TRAINING
EMSL ANALYTICAL, INC.
LABORATORY PRODUCT TRAINING

Environmental Chemistry
Chain of Custody
EMSL Order Number (Lab Use Only):

ADP# 17208

EMSL Analytical, Inc.
200 Route 130 North

Cinnaminson, NJ 08077
PHONE: 1-800-220-3675
FAX: (856) 786-5974

Report To Contact Name: James Eberts		Bill To Company: Epic Environmental Services, LLC	
Company Name: Epic Environmental Services, LLC		Attention To: James Eberts / Timothy Eberts	
Street: 80 Fork Bridge Road		Street: 80 Fork Bridge Road	
City: Pittsgrove	State/Province: NJ	Zip/Postal Code: 08318	City: Pittsgrove
Phone: 856-205-1077	Fax:	Zip/Postal Code: 08318	State/Province: NJ
Project Name: <u>Pittsgrove HS-1 ILM Additional</u>		Email Results To: <u>jeberts@epicenviro.com / teberts@epicenviro.com</u>	
U.S. State where Samples Collected: NJ		Purchase Order:	
Number of Samples in Shipment: <u>8</u>		Date of Shipment: <u>24-239c</u>	
Sample for Compliance? Yes <input checked="" type="checkbox"/> No <input type="checkbox"/> If yes, NPDES? <input type="checkbox"/> Other (Specify):			
Samples Collected by: EMSL <input type="checkbox"/> Client <input checked="" type="checkbox"/> check one		Sampled By (Signature): <u>Carey Dygala</u>	
Standard Turnaround Time: <input type="checkbox"/> 2 Weeks		The following TATs are subject to lab approval: <input type="checkbox"/> 1 Week <input type="checkbox"/> 4 Days <input type="checkbox"/> 3 Days <input type="checkbox"/> 2 Days <input type="checkbox"/> 1 Day	
Failure to complete will hinder processing of samples			
Client Sample ID	Comp	Grab	Collect Date/Time
	Matrix	Preservative	List Test(s) Needed
PHS -FB	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	4/16/25 N/A
PHS-WC-GLCK	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	0609
PHS-BF-GLCK	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	0610
PHS-WC-ADM IN	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	0615
PHS-BF-ADM IN	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	0616
Released By (Signature)		Date & Time	
<u>Carey Dygala</u>		4/16/25 0800	
Received By		Date & Time	
<u>Donnelly (w/)</u>		4/16/25 0900	
Please indicate reporting requirements: <input checked="" type="checkbox"/> Results Only <input type="checkbox"/> Results and QC <input type="checkbox"/> Reduced Deliverables <input type="checkbox"/> Disk Deliverable <input type="checkbox"/> Other			
Instructions or Comments: Note: Field pH and Field Temperature are tested on the same day as the date of sample collection.			
Analyze FIRST DRAW Samples. Analyze FLUSH samples if corresponding First Draw result is above 15.5 ppb			
(Lab) Received Temperature: _____ °C			



**Environmental Chemistry
Chain of Custody**
EMSL Order Number (Lab Use Only):

EMSL Order Number (Lab Use Only):

PHONE:
FAX:

Additional Pages of the Chain of Custody are only necessary if needed for additional sample information

[illegible]

(Lab) Received Temperature: _____°C